

# University of Kota

MBS Marg, Near Kabir Circle, KOTA (Rajasthan)-324 005, INDIA

No.F./DPAC/UOK/2016/2480

Date: 12.05.16

M/s .....

.....

## Notification of Limited Tender for the purchase of Hydrothermal autoclave reactor

**Name of Project: DST SR/SI/PC-31/2012**

**Name of PI: Dr. Neelu Chouhan**

For the purchase of Hydrothermal autoclave reactor the sealed quotations are invited on or before 20/05/2016 (2:00 pm). The tender shall be opened on 20/05/2016 (3:00 pm) at Registrar office, University of Kota, MBS Marg, Near Kabir Circle, KOTA (Rajasthan)-324 005, INDIA.

### **Specification of Hydrothermal autoclave reactor:**

S. No.	Name of Items	Specification	Quantity
1.	Hydrothermal autoclave reactor	Elegant design, , durable, easy to operate, Stainless steel structure and gas-tight, PTFE inner chamber to minimize corrosion Volume: 50 /100 mL Max Temperature: 230 °C or higher Max Pressure : 10MPa/1500psi Heating or cooling rate: 5 °C / min	01

The interested vendors submit their quotation for minimum rate with terms & conditions in closed envelope on / before above mention dates.

21/05/16  
**Registrar**  
कुल सचिव  
कोटा विश्व विद्यालय, कोटा