

# University of Kota

MBS Marg, Near Kabir Circle, KOTA (Rajasthan)-324 005, INDIA

## Additional Information for Equipment Purchase

Regarding **Tender Notice No: UOK/Chem.Deptt/14447 dated 04/10/13** for supply of Major/Minor Equipment/Instruments for the Department of Pure & Applied Chemistry, the University has modified following technical specifications for wider interest and use in academic and research:

**Modification Date:** 11-10-2013

Kindly add the following points in specifications for instruments mentioned below as:

- 1. FT-IR ATR (Sr. No. 1)**
  - Spectral resolution  $0.8\text{ cm}^{-1}$  or better
- 2. Double Beam UV-VIS Spectrometer with DRS (S. No. 2)**
  - Wavelength 190-900 or above with photomultiplier tube detector
- 3. Microwave Heating Furnace (Sr. No 7)**
  - Microwave furnace temperature up to  $1200^{\circ}\text{C}$  or more